

# Safe Cube Dispenser™ Instructions

Jan 1<sup>st</sup>, 2014

## To Set Up a Solvent Cub-a-tainer in the Safe Cube Dispenser™ Main Cabinet

1. Safe Cube Dispenser™ must be securely positioned on countertop, reference “Set-Up Guide”.
2. Open Safe Cube Dispenser™ by rotating the dispenser tray outward and down.  
Making sure the back of dispenser tray rests against front of main cabinet.
3. Turn On leak detection alarm, located on the upper-left side of the main cabinet.
4. Place a new solvent cube/container on sliding shelf inside main cabinet, with its spigot facing the front.
5. Align the cube of solvent so that the dispensing spigot is centered on the sliding shelf.
6. For 10% formalin, place a FAN-Pad GL™ in the bottom of the dispenser tray and one in the main cabinet.
7. Slide carboy to rear of main cabinet and rotate dispenser tray upward so-as-to rest on top of the carboy.

## To Dispense Solvent from the Safe Cube Dispenser™

1. Rotate dispenser tray outward and down, so back of tray rests against front of main cabinet.
2. Place the container you desire to fill against the back of dispenser tray.
3. Pull solvent cube (on sliding shelf) forward to front edge of the main cabinet.
4. Align the container to be filled directly below spigot and open spigot to fill container.
5. For smaller container a tube extension, (ie.surgical tube) on the spigot is recommended.
6. Once the container is filled, turn Off spigot and slide carboy back to the rear of the main cabinet.
7. When using a tube extension, raise the tube extension as you slide cube back into main cabinet.
8. Secure the lid on the filled container and remove filled container from the dispenser tray.
9. With filling completed and No container in dispenser tray, rotate dispenser tray upward in to close.

## Respond to Leak Detection Alarm

1. If a liquid leak or spill occurs in the main cabinet or dispenser tray, the Liquid Detection Alarm will sound.
2. Once an attendant responds to the alarm, turn Off leak detection alarm.
3. Follow your facility procedure for spill cleanup.
  - ◆ For a formalin leak, apply PolyForm-F™ and/or FAN-Pad-GL™ to absorb and neutralize the formalin.
4. Once the leak and/or spill is properly cleaned up, turn On leak detection Alarm.
5. It is recommended to test leak detection Alarm every few days, or at least once a week.

## How to Test Alarm

1. Slide Alarm out of its holder located on the outer-upper left side, turn on Alarm.
2. Leaving the Alarm ON, slide the Alarm back into its holder.
3. Wet a paper towel with tap water and place paper towel on one set of Alarm Contacts and then the other.
4. The Alarm contacts are installed in two (2) locations, the first is located inside the main cabinet in the lower-left front corner, and the second is located inside the filling tray in the lower-left front corner.
5. Once the Alarm sounds remove wet towel, SafeCubeDispenser™ is now ready for use.
6. If Alarm does not sound, replace "9-Volt" DC batteries located on the back of Alarm enclosure and re-test.
7. Slide battery cover off the of the alarm box, and replace the "9-Volt" DC battery.

Distributed By:

**NEWCOMER SUPPLY®**  
Your Histology Supplier

Middleton, WI • (800) 383 • 7799 • [www.newcomersupply.com](http://www.newcomersupply.com)