

Hematoxylin & Eosin (H&E) Progressive Stain - Technical Memo

SOLUTIONS:

	500 ml	1 Liter	1 Gallon
Hematoxylin Stain, Mayer Modified	Part 1202A	Part 1202B	Part 1202C
Eosin Y Working Solution	Part 1072A	Part 1072B	Part 1072C

Additionally Needed For H&E Staining:

Hematoxylin & Eosin (H&E) Control Slides	Part 4278		
Xylene, ACS	Part 1445		
Alcohol, Ethyl Denatured, 100%	Part 10841		
Alcohol, Ethyl Denatured, 95%	Part 10842		
Lithium Carbonate, Saturated Aqueous	Part 12215	or Scott Tap Water Substitute	Part 1380
Alcohol, Ethyl Denatured, 70%	Part 10844		

For storage requirements and expiration date refer to individual product labels.

APPLICATION:

Newcomer Supply Hematoxylin & Eosin (H&E) Progressive Stain is used for screening specimens in anatomic pathology, as well as for research, smears, touch preps and other applications. In progressive staining, tissue is left in the hematoxylin solution just long enough to reach the proper endpoint. It may be necessary to examine the slides at different timed intervals to determine when staining is optimal.

Hematoxylin Stain, Mayer Modified is a ready-to-use progressive hematoxylin that does not require filtering and does not contain chloral hydrate or alcohol. Due to the progressive staining nature of Mayer hematoxylin, over-staining is less likely and an acid alcohol differentiation step is not required in the staining process.

Eosin Y Working Solution is a ready-to-use counterstain with the ability to distinguish between the cytoplasm of different types of cells by staining cytoplasmic components differing shades and intensities of pink to red.

Quality Control: Since hematoxylin and eosin staining is the foundation of the diagnostic process, maintaining quality is of critical importance. Procedures will vary between laboratories depending upon volume of slides, automation vs manual staining, chemical hygiene and solution integrity. Longevity of hematoxylin and eosin depends upon these factors and stain quality should be regularly screened with an H&E control slide.

METHOD:

Fixation: Formalin 10%, Phosphate Buffered (1090)

Technique: Paraffin sections cut at 4 microns

Solutions: All solutions are manufactured by Newcomer Supply, Inc.

STAINING PROCEDURE:

- Deparaffinize sections thoroughly in three changes of xylene, 3 minutes each. Hydrate through two changes each of 100% and 95% ethyl alcohols, 10 dips each. Wash well with distilled water.
 - See Procedure Notes #1 and #2.
- Stain with Hematoxylin Stain, Mayer Modified, 10-20 minutes, depending on preference of nuclear stain intensity.
- Wash well in tap water for 3 minutes.
- Blue slides in Lithium Carbonate, Saturated Aqueous (12215) or Scott Tap Water Substitute (1380) for 10 dips.
 - See Procedure Note #3.
- Wash in three changes of tap water; rinse in distilled water.
- Drain water from slides; proceed to 70% ethyl alcohol for 10 dips.

- Counterstain in Eosin Y Working Solution for 30 seconds to 3 minutes, depending on preference of intensity.
- Dehydrate in two changes of 95% ethyl alcohol for 1 minute each and two changes of 100% ethyl alcohol, 10 dips each. Clear in three changes of xylene, 10 dips each; coverslip with compatible mounting medium.

RESULTS:

Nuclei	Blue
Erythrocytes and eosinophilic granules	Bright pink to red
Cytoplasm and other tissue elements	Various shades of pink

PROCEDURE NOTES:

- Drain slides after each step to prevent solution carry over.
- Do not allow sections to dry out at any point during procedure.
- Some laboratory tap water may be alkaline enough to accomplish satisfactory bluing, allowing the elimination of Step #4.
- If using a xylene substitute, follow manufacturer's recommendation for deparaffinization and clearing steps.

REFERENCES:

- Carson, Freida L. and Christa Cappellano. *Histotechnology: A Self-Instructional Text*. 5th ed. Chicago: ASCP Press, 2020. 112-115.
- Luna, Lee G. *Manual of Histologic Staining Methods of the Armed Forces Institute of Pathology*. 3rd ed. New York: Blakiston Division, McGraw-Hill, 1968. 36-38.
- Modifications developed by Newcomer Supply Laboratory.